



Aqua Compatibility

June 26th, 2023

To whom it may concern:

Please find information regarding aqua compatibility levels for our components.

	 FRESH WATER	 SEA WATER	 CHLORINE WATER
FEET			
Agilix	✓	✓	✓
DynAdapt	✓	✓	✓
Dynatrek	✓	✓	✓
Exalto	✓	✓	✓
Highlander & Highlander Max	✓	✓	✓
Pacifica & Pacifica LP	✓	✓	✓
Rush Chopart	✓	✓	✓
Rush Hipro	✓	✓	✓
Rush Kid	✓	✓	✓
Rush Rampage	✓	✓	✓
Rush Rampage LP	✓	✓	✓
Rush Rogue 2	✓	✓	✓
Rush Rover	✓	✓	✓
Shockwave	✓		
Sierra	✓	✓	✓
KNEES			
Easy Ride	✓	✓	✓
PROTEOR QUATTRO	✓	*	*
Plié 3	✓	*	*
ANKLES			
Kinnex 2.0	✓	*	*
COMPONENTS			
Connector 1D40 Titanium (Tube/Female Pyramide)	✓	✓	✓
Tube 1G01 (Aluminum Ø 30mm)	✓	✓	✓
1K202 Titanium Tibial Anchor	✓	✓	✓

* If contact with salt or chlorinated water occurs, the prosthesis must be rinsed thoroughly with fresh water immediately after contact to prevent corrosion and must be cleaned and then dried.

Please see instructions for use for full details on individual product cleaning and care. Our interactive catalog is available on our web site for further information.

Best regards,



Sophie DESBOIS

EMEAR Area - Head of marketing & customer service

20232706-UK-DIV-0011-COMPOSANTS